

## Blind sample results for pH Field Measurements/2004

Sample Identification	Measured Value	pH units Actual Value	Difference
BpH-01-1	6.84	6.89	0.05 units
BpH-01-2	4.13	4.02	0.11 units
BpH-02-1	7.44	7.42	0.02 units
BpH-02-2	7.35	7.41	0.06 units
BpH-03-1	7.42	7.40	0.02 units
BpH-03-2	3.65	4.03	0.38 units
BpH-04-1	4.16	4.02	0.14 units
BpH-04-2	6.85	6.85	0.00 units
BpH-05-1	6.80	6.88	0.08 units
BpH-05-2	3.99	4.03	0.04 units
BpH-06-1	6.80	6.86	0.06 units
BpH-06-2	7.33	7.39	0.06 units
BpH-07-1	4.00	4.02	0.02units
BpH-07-2	7.38	7.42	0.04 units
BpH-08-1	6.94	6.87	0.07 units
BpH-08-2	3.96	4.02	0.06 units
BpH-09-1	7.10	6.88	0.22 units
BpH-09-2	7.55	7.40	0.15 units
BpH-10-1	3.55	4.02	0.47* units
BpH-10-2	6.89	6.87	0.02 units
BpH-11-1	7.40	7.41	0.01 units
BpH-11-2	7.40	7.40	0.00 units
BpH-12-1	4.01	4.05	0.04 units
BpH-12-2	7.40	7.41	0.01 units

\* Value exceeded Acceptance criteria of +/- 0.4 pH units