

Blind Sample Results For pH Field Measurements -- 2006

Sample Identification	Measured Value pH Units	Actual Value pH Units	Difference
BpH0101	7.44	7.44	0.0
BpH0102	4.21	4.24	0.03
BpH0201	7.37	7.42	0.05
BpH0202	6.69	6.88	0.19
BpH0301	4.04	4.07	0.03
BpH0302	6.89	6.86	0.03
BpH0401	4.04	4.18	0.14
BpH0402	6.90	6.91	0.01
BpH0501	4.10	4.16	0.06
BpH0502	6.80	6.84	0.04
BpH0601	6.88	6.92	0.04
BpH0602	4.07	4.07	0.0
BpH0701	7.33	7.27	0.06
BpH0702	4.04	4.08	0.04
BpH0801	4.02	4.12	0.10
BpH0802	7.42	7.38	0.04
BpH0901	4.08	4.17	0.09
BpH0902	7.30	7.21	0.09
BpH1001	7.26	7.22	0.04
BpH1002	4.02	4.26	0.24
BpH1101	6.82	6.83	0.01
BpH1102	7.12	7.34	0.22
BpH1201	7.40	7.39	0.01
BpH1202	6.97	6.89	0.08